



# UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY



CC-3287

## CALIBRATION CERTIFICATE

Format No. UCCS-QF-7.8.1	Certificate No. UCCSCC20221122-017	ULR No. CC328722000005026F
Group: Pressure Indicating Devices	Discipline: Mechanical	
Customer's Name & Address	Page No.: 1 of 1	Calibration Performed: At Lab
M/s. KRM Ayurveda Pvt. Ltd.	Date of Receipt: 22-11-2022	
77, Tarun Enclave, 2nd Floor, Gate No. 2, Parwana Road, Pitam	Service Request No.: UCCSSRN20221122-001	
Pura, Delhi-110034	Certificate Issue Date: 24-11-2022	
	Date of Calibration: 23-11-2022	
	Next Due Date (Suggested by Customer): 22-11-2023	
Equipment Description (UUC) & Identification		
Name of the Instrument: Dial Pressure Gauge	Make: Duke	
(Vertical Autoclave)	Model: -----	
Range: 0 to 2.1 kg/cm <sup>2</sup>	Sr.No.: -----	
Least Count: 0.05 kg/cm <sup>2</sup>	I.D. No.: KRMAPL/AC/PG-01	
Condition of Instrument: Ok	Location: -----	
Master Equipment Used		
S.No.	Instrument Name	Range & Serial No.
1	Digital Pressure Gauge	0 To 70 bar & 53572105
Calibration Agency		Calibration Certificate No.
Transcal		TSC/21-22/14934-7
		Valid Up to
		23-01-2023
Environmental Conditions		Calibration Reference Standard
Temperature: 23 ± 1.5 °C, Humidity: 30% To 80% RH		DKD-R-6-1

### CALIBRATION RESULTS

S.No.	Set UUC Value (kg/cm <sup>2</sup> )	Average Value of Ref. Standard (kg/cm <sup>2</sup> )	Error (kg/cm <sup>2</sup> )	Expanded Uncertainty (± % of rdg.)
1	0.0	0.000	0.000	-----
2	0.5	0.489	0.011	4
3	1.0	0.982	0.018	
4	1.5	1.475	0.025	
5	2.0	1.963	0.037	
6	2.1	2.051	0.049	

\*\*\*End of Certificate\*\*\*

*Jata Shankar*  
Calibrated By

Jata Shankar  
Calibration Engineer

*Vikash Kumar*  
Reviewed/Approved By  
Vikash Kumar  
Technical Manager

Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.  
2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
3. Total liability of our Organization is limited to the invoiced amount.  
4. Calibration Certificate without signature are not valid.  
5. All our certificates are metrological traceable.  
6. This certificate relates to the item calibrated.  
7. UUC: Unit Under Calibration.  
8. Expanded Uncertainty at 95 % confidence level with coverage factor K=2.  
9. Scan this QR code to download our NABL Accreditation Scope & Certificate.



Working : W.Z.1390/Z-2, 1st Floor, Nangal Raya Extn., Pankha Road, New Delhi-110046

Regd. Office : R.Z.56A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045  
M.: 9958445875, 9313768426, 9999761885 E-Mail : uniquecalibrationservices@gmail.com



# UNIQUE CALIBRATION & CONSULTANT SERVICES

AN ISO 9001:2015 CERTIFIED CALIBRATION LABORATORY

## CALIBRATION CERTIFICATE

Format No. UCCS-QF-7.8.1		Certificate No. UCCSCCN20221122-008			
Customer's Name & Address M/s. KRM Ayurveda Pvt. Ltd. 77, Tarun Enclave, 2nd Floor, Gate No. 2, Parwana Road, Pitam Pura, Delhi-110034		Page No.: 1 of 1		Calibration Performed: At Site	
		Date of Receipt: 22-11-2022			
		Service Request No.: UCCSNSRN20221122-002			
		Certificate Issue Date: 24-11-2022			
		Date of Calibration: 23-11-2022			
		Next Due Date (Suggested by Customer): 22-11-2023			
Equipment Description (UUC) & Identification					
Name of the Instrument: Humidifier		Make: -----			
		Model: -----			
Range: As Per Instrument		Sr.No.: -----			
Least Count: -----		I.D. No.: KRMAPL/HF-01			
Condition of Instrument: Ok		Location: -----			
Master Equipment Used					
S.No.	Instrument Name	Range & Serial No.	Calibration Agency	Calibration Certificate No.	Valid Up to
1	Compact Electrical Safety Analyzer	R1/0Z01	Testncl Laboratory	TNC/JC/21-22/02/26	10-02-2023
2	Dig. Temp. & Hum. Meter With Ext. Sensor	As per manual & HTC1809CG0149	Acumen Measurements & Consultancy Pvt. Ltd.	AMC/00000398	14-02-2023
Environmental Conditions			Calibration Procedure	Calibration Reference Standard	
Temperature: 25 ± 5 °C, Humidity: 50 ± 10 %RH			UCCS 03 SOP MIS 008	-----	

### CALIBRATION RESULTS (Electrical Safety Test)

(1) Safety Analyzer Test :-					
Sr. No.	Parameter	Standard Value	Observed Value	Remarks	
1	Line to Neutral Voltage	210 - 240 V	219.9	Acceptable	
2	Line to Earth Voltage	1 - 5 V	1.52	Acceptable	
3	Load Current	0 - 2 A	0.78	Acceptable	
4	Leakage Current	0 - 500 µA	41.19	Acceptable	
5	Enclosure Current	0 - 10 µA	1.36	Acceptable	
6	Patient Leakage Current	0 - 100 µA	31.67	Acceptable	

\*\*\*End of Certificate\*\*\*

Hemraj  
Calibrated By

Hemraj

Calibration Engineer

Reviewed/Approved By  
Vikash Kumar  
Technical Manager

Note: - 1. The calibration results reported in this certificate are valid at the time of and stated condition of measurement.  
2. This report is not to be reproducing wholly or in part and cannot be used as evidence in the Court of Law and should not be used in any advertising media without our special permission in writing.  
3. Total liability of our Organization is limited to the invoiced amount.  
4. Calibration Certificate without signature are not valid.  
5. All our certificates are metrological traceable.  
6. This certificate relates to the item calibrated.  
7. UUC: Unit Under Calibration.  
8. Expanded Uncertainty at 95 % confidence level with coverage factor K=2.

Working : W.Z.1390/Z-2, 1st Floor, Nangal Raya Extn., Pankha Road, New Delhi-110046

Regd. Office : R.Z.56A, Indra Park, Gali No.25, Palam Colony, New Delhi-110045  
M.: 9958445875, 9313768426, 9999761885 E-Mail : uniquecalibrationservices@gmail.com